

WAC 296-823-18025 Protect vacuum lines. You must protect vacuum lines with liquid disinfectant traps and high-efficiency particulate air (HEPA) filters or filters of same or greater efficiency. Make sure filters are checked routinely and maintained or replaced as necessary.

[Statutory Authority: RCW 49.17.010, 49.17.040, 49.17.050. WSR 15-23-086, § 296-823-18025, filed 11/17/15, effective 12/18/15. Statutory Authority: RCW 49.17.010, 49.17.040, 49.17.050, and 49.17.060. WSR 03-09-110, § 296-823-18025, filed 4/22/03, effective 8/1/03.]