

**WAC 246-225-080 Radiographic systems other than fluoroscopic, dental intraoral, or veterinary systems—Source-to-skin or receptor distance.**

(1) *Limitation.* All radiographic systems shall be provided with a durable, securely fastened means to limit the source-to-skin distance to not less than 23 centimeters. The requirement can be met when the collimator or cone provides the required limits.

(2) *Source to receptor distance measuring device.* All radiographic systems shall be provided with a device or reference, other than a collimator light localizer, which will indicate the selected source to image receptor distance (SID) to within 2 percent of the indicated SID.

[Statutory Authority: RCW 43.70.040. WSR 91-02-049 (Order 121), recodified as § 246-225-080, filed 12/27/90, effective 1/31/91. Statutory Authority: RCW 70.98.080. WSR 83-19-050 (Order 2026), § 402-28-053, filed 9/16/83. Statutory Authority: RCW 70.98.050. WSR 81-01-011 (Order 1570), § 402-28-053, filed 12/8/80; Order 1084, § 402-28-053, filed 1/14/76. Formerly WAC 402-28-050 (part).]