

WAC 246-229-0070 What are the minimum requirements for operating and emergency procedure documentation? At a minimum, the procedures must include instruction on:

- (1) Securing the accelerator to prevent unauthorized use.
- (2) Operating the accelerator.
- (3) Responding to an emergency involving the accelerator.
- (4) Performing safety and warning device (including interlocks) checks at least every three months.
- (5) Performing radiation surveys.
- (6) Performing monitoring equipment calibration (if applicable).
- (7) Recordkeeping and/or documentation.
- (8) Bypassing safety interlocks.

[Statutory Authority: RCW 70.98.050 and [70.98.]080. WSR 02-14-050, § 246-229-0070, filed 6/27/02, effective 7/28/02.]